

CHIRANA T. Injecta, a.s. Nám. Dr. Schweitzera 194 916 01 Stará Turá Slovenská republika / Slovak Republic

EU DECLARATION OF CONFORMITY

(no. EC 24-2024)

We, CHIRANA T. Injecta a.s.,

Nám. Dr. A. Schweitzera 194, 916 01 Stará Turá, Slovakia

Single registration number (SRN): SK-MF-000003702

manufacturer of:

STERILE SINGLE USE MEDICAL DEVICES

declare in our own responsibility, that the product:

STERILE BLUNT FILL NEEDLE FOR SINGLE USE

brand name:

MEDOJECT

versions:

various sizes, with / without filter

intended purpose:

to be attached to a syringe in order to aspiration fluids from

vials or ampules during the preparation of medications

classified as Class: Is

according to classification set out in Annex VIII, Rule 2 of the EU Medical Device Regulation (MDR) 2017/745

is in compliance:

- with provisions of Annex I of the Regulation (EU) 2017/745 of the European parliament and of the council of 5th April 2017 on medical devices which apply to them
- with provisions of technical specification PS 310-0718-01, PS 310-1215-01, EN ISO 7864:2016, EN ISO 11135:2014, EN ISO 14971:2019, EN ISO 13485:2016/AC:2018, EN ISO 11607-1:2020, standards selected from the group EN ISO 14644 and EN ISO 10993, European Pharmacopoeia actual edition and other relevant ISO/EN standards mentioned in product specification and list of standards for product
- have been subject to following conformity assessment procedure laid down in the Regulation (EU) 2017/745, Article 52 (7a) relating to the Quality Management System and Technical Documentation set out in Annex IX
- are certified by the Notified Body 3EC International a.s., identification number 2265

It was issued Certificate no. 2022-MDR/QS-024/A Rev.01 dated of 11.09.2023, valid till 30.09.2027 for product. Product models are marked with CE 2265.

EU declaration of conformity is issued under the sole responsibility of the manufacturer. Signed for and on behalf of manufacturer

in Stará Turá, date 02.02.2024

CHIRANA T.Injecta, a. s.
Nám. Dr. Alberta Schweitzera 194
916 01 Stará Turá
IČO: 36794619
DIC: 2022402822
CHIRANA
IČ DPH: SK2022402822
T. Injecta

Rastislav Broska, PhD. Quality Manager



EU DECLARATION OF CONFORMITY

Product:

STERILE BLUNT FILL NEEDLE FOR SINGLE USE - MEDOJECT

Product variants:

Product REF code	Product model	Basic UDI-DI (GMN)
CH18112BF	1,2x40mm (18G x 1 1/2")	
CH18200BF	1,2x50mm (18G x 2")	858601504BFNEEDLE24SV
CH18112F	1,2x40mm with filter (18G x 1 1/2")	030001304BFNEEDLE243V
CH18200F	1,2x50mm with filter (18G x 2")	